



# Inspection Certificate

Certificate-No. 1201093



FB PAW 15.8.2 Rev. 1

Type of gauge: Diatest TD-Gauge Gr.4-DI  
Customer: GRG DIA069/11  
Identity-No. 1201093  
Standard: Diatest-Standard  
Remark: /  
Tested characteristic: Vernier depth gauge 0 – 50mm  
max. permissible error 0,05mm  
Inspection equipment: Mahr 828 PCwin-500

Measured value:			Valuation:
	10,200	10,200	ok
	20,400	20,400	ok
	30,600	30,600	ok
	40,800	40,801	ok
	50,000	50,000	ok

18.1.2012  
Date

Signature

**DIATEST**  
Hermann Költgen GmbH  
Schottener Weg 6  
D- 64289 Darmstadt

**DIATEST Hermann Költgen GmbH**  
Schottener Weg 6  
D-64289 Darmstadt / Germany  
Tel: +49 6151 - 979 0  
Fax: +49 6151 - 979 111  
<http://www.diatest.com>

This inspection certificate corresponds the requirements of the DIN EN ISO 9001:2008. Traceability to national standards is given by DKD-Calibrated set of Setting Masters.  
Gauge block DKD-K-17301-14663 and DKD-K-17301-14664.